

SOLUTION CHANGE

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

*RMJ*  
*21-3878*

Date: 02/01/2021

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:56
INTERNAL STANDARD	VERIFIED	13:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:56
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:58
BLANK TEST	0.00	13:59

