

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 01/27/2021

RM)
21-3221

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:44
INTERNAL STANDARD	VERIFIED	12:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:44
BLANK TEST	0.00	12:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:45
BLANK TEST	0.00	12:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:46
BLANK TEST	0.00	12:47

