

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/14/2021

*RMS
21-13479*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:15
INTERNAL STANDARD	VERIFIED	12:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:15
BLANK TEST	0.00	12:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:16
BLANK TEST	0.00	12:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:18
BLANK TEST	0.00	12:18

