

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 07/06/2021

Rms
21-26375-

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:06
INTERNAL STANDARD	VERIFIED	12:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:07
BLANK TEST	0.00	12:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:08
BLANK TEST	0.00	12:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:09
BLANK TEST	0.00	12:10

