

**SUPERVISOR TEST**

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*TMS*  
*21-5/11*

Date: 01/05/2021

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:57
INTERNAL STANDARD	VERIFIED	13:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:58
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:59
BLANK TEST	0.00	14:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:00
BLANK TEST	0.00	14:01

