

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 03/16/2021

*RM*  
 21-9362

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:01
INTERNAL STANDARD	VERIFIED	13:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:03
BLANK TEST	0.00	13:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:04
BLANK TEST	0.00	13:05

