

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 02/01/2021

RM
21-3878

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:57
INTERNAL STANDARD	VERIFIED	13:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:59
BLANK TEST	0.00	13:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:00
BLANK TEST	0.00	14:00

