

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 05/03/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:27
INTERNAL STANDARD	VERIFIED	12:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:27
BLANK TEST	0.00	12:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:29
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:31

IMS
22-16167

