

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
22-34821

Date: 08/30/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:25
INTERNAL STANDARD	VERIFIED	10:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:25
BLANK TEST	0.00	10:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:26
BLANK TEST	0.00	10:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:27
BLANK TEST	0.00	10:28

