

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 11/29/2022

*RMS*  
*22-18166*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:46
INTERNAL STANDARD	VERIFIED	11:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:46
BLANK TEST	0.00	11:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:47
BLANK TEST	0.00	11:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:48
BLANK TEST	0.00	11:49

