

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

*RMS
 22-703811*

Date: 08/01/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:48
INTERNAL STANDARD	VERIFIED	14:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:48
BLANK TEST	0.00	14:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:49
BLANK TEST	0.00	14:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:50
BLANK TEST	0.00	14:51

