

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 03/16/2022

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21440

*RMS*  
*22-9400*

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:20
INTERNAL STANDARD	VERIFIED	13:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:21
BLANK TEST	0.00	13:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:22
BLANK TEST	0.00	13:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:23
BLANK TEST	0.00	13:24

