

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 03/29/2022

RM>
22-11167

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21140

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:52
INTERNAL STANDARD	VERIFIED	12:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:53
BLANK TEST	0.00	12:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:54
BLANK TEST	0.00	12:55

