

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 10/25/2022

*RM*  
*22-42084*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:20
INTERNAL STANDARD	VERIFIED	09:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:21
BLANK TEST	0.00	09:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:22
BLANK TEST	0.00	09:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:23
BLANK TEST	0.00	09:24

