

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 11/15/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22260

*RMS
22-16245*

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:25
INTERNAL STANDARD	VERIFIED	12:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:26
BLANK TEST	0.00	12:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:27
BLANK TEST	0.00	12:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:28
BLANK TEST	0.00	12:29

