

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMS
22-49213

Date: 12/06/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:15
INTERNAL STANDARD	VERIFIED	14:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:17
BLANK TEST	0.00	14:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:18
BLANK TEST	0.00	14:19

