

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 02/07/2023

RMS
23-5875

OPERATOR NAME:
MICHAEL J. CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:50
INTERNAL STANDARD	VERIFIED	13:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:51
BLANK TEST	0.00	13:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:52
BLANK TEST	0.00	13:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:53
BLANK TEST	0.00	13:54

