

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
23-14755

Date: 04/05/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:51
INTERNAL STANDARD	VERIFIED	09:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:51
BLANK TEST	0.00	09:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:52
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:53
BLANK TEST	0.00	09:54

