

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

*PM*  
*23-23077*

Date: 05/23/2023

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:44
INTERNAL STANDARD	VERIFIED	09:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:45
BLANK TEST	0.00	09:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:46
BLANK TEST	0.00	09:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:47
BLANK TEST	0.00	09:48

