

SOLUTION CHANGE

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 07/03/2023

*RMS*  
*23-29869*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:18
INTERNAL STANDARD	VERIFIED	12:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:18
BLANK TEST	0.00	12:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:19
BLANK TEST	0.00	12:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:21
BLANK TEST	0.00	12:21

