

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RM5
23-36012

Date: 08/08/2023

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:48
INTERNAL STANDARD	VERIFIED	10:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:48
BLANK TEST	0.00	10:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:49
BLANK TEST	0.00	10:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:50
BLANK TEST	0.00	10:51

