

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 11/28/2023

*KMS*  
*23-53309*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:18
INTERNAL STANDARD	VERIFIED	10:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:19
BLANK TEST	0.00	10:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:20
BLANK TEST	0.00	10:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	10:21
BLANK TEST	0.00	10:22

