

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

AMS
23-49563

Date: 11/01/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23720

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:00
INTERNAL STANDARD	VERIFIED	15:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:01
BLANK TEST	0.00	15:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:02
BLANK TEST	0.00	15:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:03
BLANK TEST	0.00	15:04

