

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 02/21/2024

R.M.)
24-7700

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:27
INTERNAL STANDARD	VERIFIED	07:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:29
BLANK TEST	0.00	07:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:30
BLANK TEST	0.00	07:31

