

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*AMS*  
*2/1-9/02*

Date: 03/01/2024

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:05
INTERNAL STANDARD	VERIFIED	14:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:05
BLANK TEST	0.00	14:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:07
BLANK TEST	0.00	14:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:08
BLANK TEST	0.00	14:09

