

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 03/02/2021

RMJ
21-7559

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:06
INTERNAL STANDARD	VERIFIED	13:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:06
BLANK TEST	0.00	13:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:07
BLANK TEST	0.00	13:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:08
BLANK TEST	0.00	13:09

