

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RMS
21-16287

Date: 05/03/2021

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:53
INTERNAL STANDARD	VERIFIED	13:53
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	13:53
BLANK TEST	0.00	13:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:54
BLANK TEST	0.00	13:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:55
BLANK TEST	0.00	13:56

