

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 02/23/2021

*RMS*  
*21-6632*

OPERATOR NAME:  
 MIOCHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:11
INTERNAL STANDARD	VERIFIED	12:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:12
BLANK TEST	0.00	12:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:13
BLANK TEST	0.00	12:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:14
BLANK TEST	0.00	12:15

