

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 02/09/2021

*RM >
21-4/9/10*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:27
INTERNAL STANDARD	VERIFIED	12:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:27
BLANK TEST	0.00	12:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:28
BLANK TEST	0.00	12:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:30

