

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 03/23/2021

RM
21-10253

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20050

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:16
INTERNAL STANDARD	VERIFIED	13:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:16
BLANK TEST	0.00	13:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:18
BLANK TEST	0.00	13:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:19
BLANK TEST	0.00	13:19

