

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 03/30/2021

*ZMS*  
*21-11281*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:12
INTERNAL STANDARD	VERIFIED	12:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:13
BLANK TEST	0.00	12:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:14
BLANK TEST	0.00	12:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:15
BLANK TEST	0.00	12:16

