

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 05/04/2021

RMS
21-16597

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:27
INTERNAL STANDARD	VERIFIED	15:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:27
BLANK TEST	0.00	15:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:28
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:30
BLANK TEST	0.00	15:30

