

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 09/01/2021

*RMJ*  
*21-344/34*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:09
INTERNAL STANDARD	VERIFIED	14:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:09
BLANK TEST	0.00	14:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:10
BLANK TEST	0.00	14:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:11
BLANK TEST	0.00	14:12

