

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 06/03/2022

RM)
22-2110

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:08
INTERNAL STANDARD	VERIFIED	11:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:08
BLANK TEST	0.00	11:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:09
BLANK TEST	0.00	11:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:10
BLANK TEST	0.00	11:11

