

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 07/01/2022

*RMS*  
*22-25191*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:26
INTERNAL STANDARD	VERIFIED	13:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:27
BLANK TEST	0.00	13:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:28
BLANK TEST	0.00	13:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:29
BLANK TEST	0.00	13:30

