

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

*RMS*

Date: 08/01/2022

*72-30384*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:47
INTERNAL STANDARD	VERIFIED	14:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:47
BLANK TEST	0.00	14:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:49
BLANK TEST	0.00	14:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:50
BLANK TEST	0.00	14:50

