

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RMS
22-8373

Date: 03/08/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:04
INTERNAL STANDARD	VERIFIED	13:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:04
BLANK TEST	0.00	13:05
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:05
BLANK TEST	0.00	13:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:07
BLANK TEST	0.00	13:07

