

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

RM

Date: 08/11/2022

22-32549

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 84622
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:16
INTERNAL STANDARD	VERIFIED	09:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:16
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:17
BLANK TEST	0.00	09:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:18
BLANK TEST	0.00	09:19

