

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 01/03/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21390

ZMS
22-296

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:30
INTERNAL STANDARD	VERIFIED	13:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:30
BLANK TEST	0.00	13:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:31
BLANK TEST	0.00	13:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:32
BLANK TEST	0.00	13:33

