

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

*RMS*  
*22-435*

Date: 01/04/2022

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 213390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:16
INTERNAL STANDARD	VERIFIED	14:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:16
BLANK TEST	0.00	14:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:17
BLANK TEST	0.00	14:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:18
BLANK TEST	0.00	14:19

