

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

RMS
22-30705

Date: 08/03/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:39
INTERNAL STANDARD	VERIFIED	13:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:40
BLANK TEST	0.00	13:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:41
BLANK TEST	0.00	13:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:42
BLANK TEST	0.00	13:43

