

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 09/06/2022

RMS
22-35906

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:56
INTERNAL STANDARD	VERIFIED	13:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:56
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:58
BLANK TEST	0.00	13:59

