

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

2ms

Date: 02/01/2023

23-4901

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:13
INTERNAL STANDARD	VERIFIED	14:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:13
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:14
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:15
BLANK TEST	0.00	14:16

