

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

2ms
23-134/58

Date: 03/28/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:50
INTERNAL STANDARD	VERIFIED	10:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:51
BLANK TEST	0.00	10:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:53
BLANK TEST	0.00	10:54

