

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 09/19/2023

*RM*  
*23-42835*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:39
INTERNAL STANDARD	VERIFIED	10:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:39
BLANK TEST	0.00	10:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:40
BLANK TEST	0.00	10:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:41
BLANK TEST	0.00	10:42

