

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 09/26/2023

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23190

LMJ
23-44115

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:02
INTERNAL STANDARD	VERIFIED	12:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:02
BLANK TEST	0.00	12:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:03
BLANK TEST	0.00	12:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:05
BLANK TEST	0.00	12:05

