

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 01/02/2023

RMS
23-326

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:05
INTERNAL STANDARD	VERIFIED	14:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:05
BLANK TEST	0.00	14:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:06
BLANK TEST	0.00	14:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:07
BLANK TEST	0.00	14:08

