

SOLUTION CHANGE

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 09/01/2023

*RM >*  
*23-39985*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:06
INTERNAL STANDARD	VERIFIED	13:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:07
BLANK TEST	0.00	13:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:08
BLANK TEST	0.00	13:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:09
BLANK TEST	0.00	13:10

