

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMS
23-51235

Date: 11/14/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:54
INTERNAL STANDARD	VERIFIED	11:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:54
BLANK TEST	0.00	11:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:56
BLANK TEST	0.00	11:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:57
BLANK TEST	0.00	11:58

