

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 03/01/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23430

RMS
24-9/02

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:00
INTERNAL STANDARD	VERIFIED	14:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:01
BLANK TEST	0.00	14:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:02
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:03
BLANK TEST	0.00	14:04

